

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
CAB number 2362, Calibration laboratory  
Raisova 547, Holínské Předměstí, 506 01 Jičín

**CMC for the field of measured quantity: Pressure**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Deformation and digital manometers, pressure transducers and pressure measuring chains	-95 kPa	to	-7 kPa	relative pressure	gas	0.04 %	Comparison with a standard calibrator	KL-PM-0101	
		-7 kPa	to	14 kPa			0.0028 kPa			
		14 kPa	to	14 MPa			0.02 %			
		14 MPa	to	60 MPa	relative pressure	liquid	0.03 %	Comparison with a piston manometer		
		60 MPa	to	70 MPa			0.1 %	Comparison with a digital manometer		
		5 kPa	to	70 kPa	absolute pressure	gas	0.028 kPa	Comparison with a standard calibrator		
		70 kPa	to	14 MPa			0.02 % + 0.014 kPa			
		14 MPa	to	60 MPa	absolute pressure	liquid	0.03 % + 0.1 kPa	Comparison with a piston manometer		
		60 MPa	to	70 MPa			0.10 % + 0.1 kPa	Comparison with a digital manometer		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
CAB number 2362, Calibration laboratory  
Raisova 547, Holínské Předměstí, 506 01 Jičín

**CMC for the field of measured quantity: Temperature**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
1*	Resistance temperature sensors (with/without a transducer), direct indicating thermometers and measuring chains with resistance temperature sensors					Comparison with a Pt100 standard in liquid baths and dry block calibrators	KL-PM-0001	
2*	Thermoelectric temperature sensors (with/without a transducer), direct indicating thermometers and measuring chains with thermocouples					Comparison with a Pt100 standard in liquid baths and dry block calibrators	KL-PM-0002	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
CAB number 2362, Calibration laboratory  
Raisova 547, Holínské Předměstí, 506 01 Jičín

Ord. numb er <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Loca- tion
		min.	unit					
3*	Dial thermometers					Comparison with a Pt100 standard in liquid baths and dry block calibrators	KL-PM-0005	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

JSP, s.r.o.  
CAB number 2362, Calibration laboratory  
Raisova 547, Holínské Předměstí, 506 01 Jičín

CMC for the field of measured quantity: Air humidity

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min	unit	max	unit					
1*	Relative humidity / hygrometers and measuring chains incl. humidity probes	5 % RH	to	30 % RH		Air Temperature (7 to 60) °C	1.2 % RH	Comparison with a reference hygrometer	KL-PM-0201	
		30 % RH	to	50 % RH			1.3 % RH			
		50 % RH	to	70 % RH			1.4 % RH			
		70 % RH	to	80 % RH			1.5 % RH			
		80 % RH	to	90 % RH			1.6 % RH			
		90 % RH	to	95 % RH			1.8 % RH			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



-3-

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

JSP, s.r.o.  
CAB number 2362, Calibration laboratory  
Raisova 547, Holínské Předměstí, 506 01 Jičín

CMC for the field of measured quantity: Electrical quantities

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min	unit					
1*	Measurement and simulation of temperature sensor signals (resistance temperature sensors, thermocouple temperature sensors)	0 Ω	to	600 Ω	0.007 % + 3 mΩ 0.007 % + 30 mΩ 0.004 % + 1.7 µV	Comparison with a reference multimeter	KL-PM-0006	
		600 Ω	to	6,000 Ω				
	-10 mV to 100 mV							
	Measurement and simulation of unified output signals	0 V	to	10 V	0.0035 % + 47 µV	Comparison with a reference multimeter		
		0 mA	to	20 mA	0.005 % + 0.0008 mA	Indirect current measurement		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

*"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and for the certificate itself."*

